



CERTIFICATE OF ACCREDITATION

ANSI-ASQ National Accreditation Board

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

MadgeTech, Inc.

6 Warner Road

Warner, NH 03278

has been assessed by ANAB
and meets the requirements of international standard

ISO/IEC 17025:2005

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations to which this accreditation applies.

AC-2481

Certificate Number


ANAB Approval

Certificate Valid: 05/08/2018-06/28/2019
Version No. 002 Issued: 05/08/2018



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005 AND ANSI/NCSL Z540-1-1994 (R2002)

MadgeTech, Inc.
6 Warner Road
Warner, NH 03278
Dianne Moulton (603) 746-2011
dianne@madgetech.com

CALIBRATION

Valid to: June 28, 2019

Certificate Number: AC-2481

Thermodynamic

Table with 4 columns: Parameter/Equipment, Range, Expanded Uncertainty of Measurement (+/-), Reference Standard, Method, and/or Equipment. It lists four rows of resistance thermometry data.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

- 1. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2481.

Handwritten signature of R. D. [unclear]
Vice President