

Firmware Upgrade for Interface Cables and Transceivers

To run the firmware upgrade for a MadgeTech interface cable or transceiver, check to ensure that the appropriate upgrade file(s) have been downloaded. The computer that will run the upgrade will need to have the latest MadgeTech 4 data logger software already installed.

Required before the upgrade:

- Access to a computer with:
 - MadgeTech 4 Data Logger Software installed
 - The firmware upgrade zip file for the device found at: <http://www.madgetech.com/firmware>

Upgrading your device

1. Download the correct firmware file and save it to the computer.
2. Launch the MadgeTech software.
3. Connect the interface cable or transceiver to the PC.
4. In the MadgeTech Software choose **File > Options** (*shown right*)
5. Navigate to the Communication tab on the left column.
6. In the Detected interfaces box, locate the interface cable or transceiver and select **Update**.
7. A prompt will appear, saying that the firmware will be updated on the device and to keep the device plugged in, select **OK**.
8. Browse to the file that was downloaded in step 1 and click **Open**.
9. Wait for the firmware upgrade process to complete.
10. Repeat these steps for each interface cable or transceiver that needs to be updated.
11. Once the process is complete, quit and restart the MadgeTech Software.

MadgeTech, Inc.
(603) 456-2011 Phone
(603) 456-2012 Fax
www.madgetech.com
support@madgetech.com

6 Warner Road
Warner NH 03278

