

CERTIFICATE OF COMPLIANCE

Certificate Number: SGSNA/20/SUW/00042X

Contract Number: 800740
Certificate Project Number: SUW-CERT191200505



Certified Product: Temperature and Humidity Data Logger
Trademarks: -
Model(s): 902213-00 RHTemp1000IS, 902214-00 Temp1000IS-1, 902215-00 Temp1000IS-2, 902216-00 Temp1000IS-5.25, 902217-00 Temp1000IS-7, 902218-00 RHTEMP1000IS-KR, 902219-00 Temp1000IS-1-KR, 902220-00 Temp1000IS-2-KR, 902221-00 Temp1000IS-5.25-KR, 902222-00 Temp1000IS-7-KR

Technical Data: 3.6Vdc, 1Ah
Class I, Division 1, Groups A-D, -40°C<Tamb< +80°C, T4A and
Class I, Division 2, Groups A-D, -40°C<Tamb< +80°C, T4A

Certificate Holder: MadgeTech, Inc.
6 Warner Road, 03278, Warner, New Hampshire, United States of America

This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at www.sgs.com or using the QR code below. The product is certified according to ISO/IEC Guide 17067, Conformity assessment - Fundamentals of product certification, System 3, and in accordance with:

FM Class 3600, Rev. Jan. 2018
FM Class 3610, Rev. Jan. 2018
CAN/CSA C22.2 No. 60079-0:15 Rev. October 2015
CAN/CSA C22.2 No. 60079-11:14

Authorized by:

Effective date: 14 May 2020



Paul Krauss
Certifier



Certification Body

Consumer and Retail Services, a division of SGS North America Inc.
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA
t +1 770 570 1800 f +1 770 277 1240 www.sgs.com

