



CERTIFICATE OF COMPLIANCE

Certificate Number: SGSNA/20/SUW/00042X

Contract Number: 800740

Certificate Project Number: SUW-CERT191200505

Certified Product: Temperature and Humidity Data Logger

Trademarks: -

Model(s): 902213-00 RHTemp1000IS, 902214-00 Temp1000IS-1, 902215-00 Temp1000IS-2, 902216-00 Temp1000IS-5.25, 902217-00 Temp1000IS-7, 902218-00 RHTEMP1000IS-KR, 902219-00 Temp1000IS-1-KR, 902220-00 Temp1000IS-2-KR, 902221-00 Temp1000IS-5.25-KR, 902222-00 Temp1000IS-7-KR

Technical Data: 3.6Vdc, 1Ah

Class I, II, III, Division 1, Groups A-G, -40°C<Tamb< +80°C, T4A and
Class I, II, III, Division 2, Groups A-D, F, G, -40°C<Tamb< +80°C, T4A

Certificate Holder: MadgeTech, Inc.

6 Warner Road, 03278, Warner, New Hampshire, United States of America

This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at www.sgs.com or using the QR code below. The product is certified according to ISO/IEC Guide 17067, Conformity assessment - Fundamentals of product certification, System 3, and in accordance with:

FM Class 3600, Rev. Jan. 2018

FM 3610, Jan. 2018

CAN/CSA C22.2 No. 60079-0:15 Rev. October 2015

CAN/CSA C22.2 No. 60079-11:14

Conditions of Acceptability:

This product is to be used with Tadiran TL-2150 1/2AA Type Batteries only.

Authorized by:

Effective date: 17 June 2021



Paul Krauss
Certifier



Certification Body

Consumer and Retail Services, a division of SGS North America Inc.
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA
t +1 770 570 1800 f +1 770 277 1240 www.sgs.com

